

# Biological Safety Training

Thierry Savin



October 2019



UNIVERSITY OF  
CAMBRIDGE

Department of Engineering



## General philosophy:

- ▶ Take **reasonable** care of your own safety and **others'**
- ▶ Use equipments, chemicals, bio-agents **safely**
- ▶ Only undertake tasks for which adequate **training** has been given
- ▶ **Cooperate** with colleagues to ensure a healthy and safe workplace
- ▶ Follow the University and Department's **procedures**
- ▶ **Report** all incidents and hazards



**HSE**  
Health & Safety  
Executive

<http://www.hse.gov.uk/>

- ▶ Responsible for the *encouragement, regulation and enforcement of workplace health, safety and welfare*
- ▶ Get advice from advisory committees  
e.g. *Advisory Committee on Dangerous Pathogens (ACDP)*
- ▶ Provides specific guidelines
- ▶ Numerous publications on the website



## Central Administration

- ▶ Occupational Health and Safety Service (OHSS)
- ▶ Safety Office
  - ▶ Publications, links:  
<http://www.admin.cam.ac.uk/offices/safety/>
  - ▶ Training:  
<http://www.training.cam.ac.uk/ohss/>

## Department of Engineering

- ▶ Department of Engineering Health & Safety:  
<http://safety.eng.cam.ac.uk/>
- ▶ Safety Committee & Safety Officer
- ▶ Local Safety Coordinators (LSC)
- ▶ Laser, Radiation & *Biological* Safety Officers



## Main functions:

- ▶ Review & approve<sup>1</sup> biological Risk Assessments
- ▶ Provide specialist advice on biological safety issues
  - ▶ containment, storage, security, transport, disposal, disinfection, emergency. . .
- ▶ Provide and arrange training for handling bio-agents
- ▶ Ensure biological incidents are reported and investigated

⇒ You have the main responsibility!

---

<sup>1</sup>most of the time. . .



~~Don't do it!~~ How can you do it safely?

**Hazard:** something with the potential to cause harm

- ▶ Physical, chemical, ergonomic, *biological*

**Risk:** likelihood of the potential harm from hazard being realised

- ▶ likelihood of occurrence
- ▶ potential severity of the outcome
- ▶ number of people who may be affected
- ▶ type of people who may be affected


⇒ Control strategies



### Control of Substances Hazardous to Health (CoSHH)

- ▶ What is the substance?
  - ▶ Chemicals, dusts, *biological agents*, other
- ▶ How will you be using it
- ▶ How might exposure occur
- ▶ Who may be exposed
- ▶ For how long
- ▶ How to dispose of it

⇒ Control strategies







HSE define a biological agent as “a micro-organism, cell culture, or human endoparasite which may cause infection, allergy, toxicity or other hazard to human health”

- ▶ micro-organisms: bacteria, viruses, moulds...

Routes of infection:

- ▶ ingestion (mouth)
- ▶ inhalation via aerosols (respiratory tract)
- ▶ instillation (eyes mucosa)
- ▶ percutaneous (damaged skin)

In general, no exposure standards are set for biological agents  
(no dose-response relationship...)



When classifying a biological agent it should be assigned to one of the following groups according to its level of risk of infection to humans:

hazard group	cause human <sup>2</sup> disease	spread to community	prophylaxis or treatment	examples
1	unlikely	-	-	disabled <i>E. Coli</i>
2	potential	unlikely	available	<i>Streptococcus</i>
3	severe	potential	available	<i>Hepatitis B</i>
4	severe	likely	not available	<i>Ebola</i>

<sup>2</sup>healthy, no pregnancy, no allergy...



Choosing the containment level:

hazard group (pathogens)	containment level (laboratories)
1	1
2	2
3	3
4	4

### Cell lines:

Hazard	Cell line	Containment level
Low	Well characterised, authenticated cell lines of human or primate origin; low risk of endogenous infection with a biological agent presenting no apparent harm to laboratory; tested for the most serious pathogens.	CL1
Medium	Finite or continuous cell lines/strains of human or primate origin; not fully characterised or authenticated, except where there is a high risk of endogenous biological agents, e.g. blood-borne viruses.	CL2
High	Cell lines with endogenous biological agents or cells that have been deliberately infected.	Appropriate to the agent
	Primary cells from blood or lymphoid cells of human or simian origin.	Appropriate to the potential risk



### Live animals

- ▶ No animal facility in the department

### Human tissue (surgical specimens or cadavers)

- ▶ Ethical approval required
- ▶ The department does not have permission to **store** human tissue samples under the Human Tissue Act 2004



### Risk assessment (RA) includes biological agents

1. You must fill RA form & CoSHH forms<sup>3</sup>
2. You and your supervisor must sign all forms
3. Bring all forms to Biological Safety Officer who will review them and sign the RA form upon approval
4. Bring all forms to Local Safety Coordinator who will sign all upon approval
5. Bring all forms to Departmental Safety Officer who will sign all upon approval

---

<sup>3</sup><http://safety.eng.cam.ac.uk/procedures/riskassessment>



- ▶ GMO: organisms whose genes have been **artificially** altered to modify their characteristics
- ▶ Controls are set out in the “*Genetically Modified Organisms (Contained Use) Regulations 2000*”
- ▶ “Contained Use”: limit contact between GMO and **people** or the **environment**. It relates to:
  - ▶ the actual process of genetic modification
  - ▶ the use, storage, transport & destruction of GMOs
- ▶ Need a **specific, separate** Risk Assessment made under the GMO (CU) Regulations 2000



- ▶ For GM micro-organisms, the risk assessment **must** contain details about:
  - ▶ **recipient micro-organism** (hazard group, wild/disable, survival)
  - ▶ **inserted gene** (expression, action)
  - ▶ **donor** (species)
  - ▶ **vector** (viral, cellular, plasmid)
  - ▶ **resulting GM micro-organisms**
- ▶ Assess separately the risks for **human health & safety** and for **environmental harm**.
- ▶ Assign a **provisional** class & containment level:

Risk	Class (GM micro-organism)	Containment level (laboratories)
Negligible	1	1
Low	2	2
Moderate	3	3
High	4	4





- ▶ GMO Risk Assessment must be signed by:
  - ▶ You and your supervisor
  - ▶ BSO
  - ▶ Head of Department
- ▶ Only **Class 1** work is allowed in the department
- ▶ We encourage users/supervisors to attend the training  
<http://www.training.cam.ac.uk/ohss/>  
search "Risk Assessment of Genetically Modified Organisms"  
(next one: Dec 3<sup>rd</sup>)
- ▶ HSE guidance can be found here:  
<http://www.hse.gov.uk/biosafety/gmo/acgm/acgmcomp/>




## Hygiene

- ▶ No eating, no drinking, no applying cosmetic, no smoking. . .
- ▶ Always wear lab coat
- ▶ Avoid contamination of keyboards, telephone, door handles. . .
- ▶ Wash your hand before leaving the laboratory
- ▶ Remove laboratory coat before leaving (hung in the Biolab foyer)



### Housekeeping

- ▶ Keep floor and benches **tidy**
- ▶ Keep floor and benches **clean** and **disinfected**
- ▶ All containers clearly **labeled**: content, user, date
- ▶ Do not leave **unattended scalpels** on benches; use sharp bins
- ▶ Waste container closed and marked
- ▶ Do not leave **used glassware** in the sinks: put it in the dishwasher (will be started by the technician once full)
- ▶ Some items need to be washed by hand: do it before you leave
- ▶ **Clean** balances and **turn off** after use





## Instruments

- ▶ Only use instruments you've been trained for

Ask the technician

- ▶ If you find the instrument broken

Tell the technician immediately

- ▶ If the instrument breaks

Tell the technician immediately

Alex Casabuena at Nanoscience on **wednesday mornings**



### Equipment

- ▶ Impervious & resistant surfaces, easy to clean
- ▶ Autoclave on site

### System of work

- ▶ No eating, drinking, smoking, etc.
- ▶ Door closed during work
- ▶ Observation window
- ▶ Protective clothing
- ▶ Disinfection available
- ▶ Minimize aerosol production

### Waste

- ▶ Dedicated bins for contaminated solid waste
- ▶ Validated inactivation of liquid waste



### Equipment

- ▶ Impervious & resistant surfaces, easy to clean
- ▶ Autoclave **in the building**
- ▶ **Negative pressure if possible, microbiological safety cabinet**

### System of work

- ▶ No eating, drinking, smoking, etc.
- ▶ Door closed during work, **access restricted**
- ▶ Observation window, **signage**
- ▶ **Side or back fastening lab coats, appropriate gloves, spill trays**
- ▶ Disinfection available, **specified decontamination procedures**
- ▶ Minimize aerosol production, **control aerosol dissemination**
- ▶ **Safe storage of biological agents**
- ▶ **Written record of staff training**

### Waste

- ▶ Dedicated bins for contaminated solid waste
- ▶ Validated inactivation of liquid waste

# Containment Levels


## Microbiological Safety Cabinets



UNIVERSITY OF  
CAMBRIDGE

Department of Engineering

- ▶ Different from a fume hood (for hazardous chemicals)
- ▶ Requirement according to risk assessment (e.g. aerosols)
- ▶ Class II cabinet with HEPA<sup>4</sup> filter protects **worker** and **work**



credit: [nuaire.com](http://nuaire.com)

<sup>4</sup>High-Efficiency Particulate Air

# Containment Levels


Department of Engineering



UNIVERSITY OF  
CAMBRIDGE

Department of Engineering

- ▶ INO-03 to INO-07 Biolab 1 - Inglis mezzanine: **CL1+CL2**
- ▶ ING-09/10 Biolab 2 - *via* materials teaching lab: **CL1**
- ▶ IN1-06 EfLS teaching lab: **CL1**
- ▶ ISG-87 Geotech lab: **CL1**





# Containment Levels

Nanoscience Centre



UNIVERSITY OF  
CAMBRIDGE

Department of Engineering

- ▶ Bay 4: CL1, will host a CL2 space in the very near future
- ▶ Module 1: CL1



### Contaminated solid waste

- ▶ Dispose into **dedicated** clinical waste bins
- ▶ Do not tape clinical waste to benches
- ▶ Do not overfill
- ▶ Company (Novus Environmental) pick-up bins at the beginning of the month (first thursday)



credit: [initial.co.uk](http://initial.co.uk)

### Contaminated sharps

- ▶ Must be placed in dedicated sharps bins
- ▶ Place pipette tips in provided **puncture-proof** containers
- ▶ Orange-lidded bins for blood (human or animal) contamination



credit: [steroplast.co.uk](http://steroplast.co.uk)



Sharps can be:

- ▶ Needles
- ▶ Blades eg scalpels, microtome blades
- ▶ Other (medical) instruments that could cause an injury by cutting/pricking the skin

Items such as glass coverslips, microscope slides, glass Pasteur pipettes can also be 'classed' as sharps, as can bone fragments.

**DO NOT place any sharps into the general waste stream**

Appropriate secure sharps disposal containers must be used and located close to the work area at a suitable height.

Safe use of sharps:

<https://www.safety.admin.cam.ac.uk/system/files/hsd192b.pdf>



### Biological liquid waste

- ▶ Must be **deactivated**
  - ▶ **Treat** with an appropriate disinfectant (e.g. Distel)
  - ▶ Allow sufficient **contact time** (see manufacturer's guidelines)
- ▶ **Dispose** in an appropriate lab sink
- ▶ **Flush down** with copious amounts of water



Transport of hazardous substances should be **minimized**

- ▶ **Internal transport (same building)**
  - ▶ Minimum quantities
  - ▶ Secondary containment
  - ▶ Use (goods) lift if possible
- ▶ **Transport between buildings**
  - ▶ Sealed, leakproof primary container
  - ▶ Leakproof, closable secondary container
  - ▶ signage, list of content
- ▶ **External transport**
  - ▶ Highly regulated, extremely complicated
  - ▶ Talk to BSO



### General procedure

- ▶ Evacuate area (if you can not assess the spill, large spills)
- ▶ BSO must **assess** the spill
- ▶ Ventilate area, allow aerosols to settle
- ▶ **Contain**, and apply **disinfectant** (e.g. Virkon powder)
- ▶ **Clean** up, fumigation may be needed
- ▶ **Waste disposal**
- ▶ Need **permission** for re-entry



- ▶ **Bleed** the wound and **wash** the area with soap and water
- ▶ Scrubbing should be avoided
- ▶ **Cover** the wound with a waterproof dressing
- ▶ Contamination on skin eye or mucous membranes should be washed **immediately**
- ▶ **Seek advice** on post-exposure prophylaxis
- ▶ **Report** the accident

# Risk Assessments & Enquiries

Biological Safety in the Department of Engineering



UNIVERSITY OF  
CAMBRIDGE  
Department of Engineering

BSO Thierry Savin

- ▶ **email** [ts573@cam.ac.uk](mailto:ts573@cam.ac.uk)
- ▶ **phone** x32762
- ▶ **office** BE3-19

LSC (Engineering) Athina Markaki

- ▶ **email** [am253@cam.ac.uk](mailto:am253@cam.ac.uk)
- ▶ **phone** x66417
- ▶ **office** BE3-21

LSC (Nanoscience) David Hasko

- ▶ **email** [dgh4@cam.ac.uk](mailto:dgh4@cam.ac.uk)
- ▶ **phone** x37496

LSC (Geotech) Giovanna Biscontin

- ▶ **email** [gb479@cam.ac.uk](mailto:gb479@cam.ac.uk)
- ▶ **phone** x68044

DSO Gary Steele

- ▶ **email** [safety-office@eng.cam.ac.uk](mailto:safety-office@eng.cam.ac.uk)
- ▶ **phone** x32740
- ▶ **office** INO-18

Biotechnician Alex Casabuena

- ▶ **email** [ac958@cam.ac.uk](mailto:ac958@cam.ac.uk)
- ▶ **phone** x48542
- ▶ **office** BCO-10

## References

- ▶ Safe biological practice at the University of Cambridge:  
<http://www.safety.admin.cam.ac.uk/publications>
- ▶ Biological Safety @CUED:  
<http://safety.eng.cam.ac.uk/procedures/biological>
- ▶ The Approved List of biological agents:  
<http://www.hse.gov.uk/pubns/misc208.pdf>